



## *Certificate of Conformance*

# **RAPICIDE TEST STRIP**

**Part Number:** ML02-0120

**Lot Number:** 025034

**Expiration Date:** 2016-01

I certify that the Rapicide test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the Rapicide test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

A handwritten signature in black ink, appearing to read "Doris J." followed by a stylized "J".

Quality Analyst

03FEB2014

Analysis Date

A handwritten signature in blue ink, appearing to read "Sue Wahr".

Quality Assurance

03FEB2014

Release Date

**MINNTECH CORPORATION**

14605 28th Avenue North

Plymouth, Minnesota 55447

Telephone: (763) 553-3300

Fax (763) 553-3387