

Certificate of Conformance

McKESSON[®] OPA/28 Test Strips

Part Number: **73-OPA28B**

Lot Number: 025046

Vial Lot Number: 032046

Expiration Date: 2017-12-31

Quality Control Report:

MEDIVATORS certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of $0.50\% \pm 0.05\%$ OPA.

MEDIVATORS certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of $0.35\% \pm 0.05\%$ OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.



Quality Assurance

5/18/2016

Release Date

Medivators Inc.
14605 28th Avenue N
Minneapolis, MN 55447
Tel: 800.444.4729
Fax: 763.553.3387
www.medivators.com

McKESSON[®] is a registered trademark of McKesson Corporation.