



Certificate of Conformance

RAPICIDE® Test Strip

Part Number:	ML02-0120
Lot Number:	025610
Expiration Date:	2016-03

I certify that the RAPICIDE test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the RAPICIDE test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

A handwritten signature in black ink that appears to read "Sami Walsh".

Quality Analyst

13May2014

Analysis Date

A handwritten signature in black ink that appears to read "Pam B. Bales".

Quality Assurance

13May2014

Release Date

Medivators Inc.

14605 28th Avenue North
Minneapolis, MN 55447 USA

Telephone: (763) 553-3300

Fax (763) 553-3387

www.medivators.com