



Certificate of Conformance

RAPICIDE® Test Strip

Part Number: ML02-0120

Lot Number: 025671

Expiration Date: 2016-05

I certify that the RAPICIDE test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the RAPICIDE test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.


Linda Heidecker
Quality Analyst

14Jul2014
Analysis Date


David Walker
Quality Assurance

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Release Date 14Jul2014 Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com