



Certificate of Conformance

RAPICIDE® Test Strip

Part Number: ML02-0120

Lot Number: 026323

Expiration Date: 2016-10

I certify that the RAPICIDE test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the RAPICIDE test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.



Quality Analyst

20Nov2014
Analysis Date



Quality Assurance

20Nov2014
Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com