

Certificate of Analysis

RAPICIDE[®] PA

Part A

Product Code: ML02-115

Lot Number: 026518

Expiration Date: 2020/02

Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0% (w/w)	<u>22.0 %</u>
PAA	4.0 – 6.0 % (w/w)	<u>5.2 %</u>
Acetic Acid	8.0 – 10.0% (w/w)	<u>9.1 %</u>
pH, (as is)	0.5 – 1.1 (as is)	<u>0.7</u>


This lot has been reviewed and meets specified requirements.



Quality Analyst

10 August 2018

Analysis Date



Quality Assurance

20-AUG-18

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE[®] is a registered trademark of Medivators Inc.