



Certificate of Conformance

RAPICIDE® Test Strip

Part Number: ML02-0120


Lot Number: 027629

Expiration Date: 2017-06

I certify that the RAPICIDE test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the RAPICIDE test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.



Quality Analyst

05Aug2015

Analysis Date



Quality Assurance

05Aug2015

Release Date

Medivators Inc.

14605 28th Avenue North

Minneapolis, MN 55447 USA

Telephone: (763) 553-3300

Fax (763) 553-3387

www.medivators.com