



## Certificate of Analysis

# RAPICIDE® PA

## Part B

Product Code: ML02-0115  
Lot Number: 027916  
Expiration Date: 2018/05

Components	Specification	Analysis
Sodium	743.0 - 909.0 mEq/L	784.2 mEq/L
Benzotriazole	0.9 – 1.1% w/w	0.9 %w/w
pH	11.4 to 12.4 as is	12.0

This lot has been reviewed and meets specified requirements.

  
Quality Analyst

01 Dec 2016  
Analysis Date

  
Quality Assurance 

05 Dec 2016  
Release Date

Medivators Inc.  
14605 28th Avenue North  
Plymouth, Minnesota 55447  
Telephone: (763) 553-3300  
Fax (763) 553-3387

RAPICIDE® is a registered trademark of Medivators Inc.