



## Certificate of Conformance

# **RAPICIDE® Test Strip**

**Part Number:**

ML02-0120

**Lot Number:**

028019

**Expiration Date:**

2017-09

I certify that the RAPICIDE test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the RAPICIDE test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

Quality Analyst

22Oct2015

Analysis Date

Quality Assurance

22Oct2015

Release Date

**Medivators Inc.**

14605 28th Avenue North

Minneapolis, MN 55447 USA

Telephone: (763) 553-3300

Fax (763) 553-3387

[www.medivators.com](http://www.medivators.com)