



## Certificate of Conformance

### **RAPICIDE® Test Strip**

**Part Number:** ML02-0120

**Lot Number:** 028505

**Expiration Date:** 2017-11

I certify that the RAPICIDE test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the RAPICIDE test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

  
\_\_\_\_\_  
Quality Analyst

22Jan2016  
Analysis Date

  
\_\_\_\_\_  
Quality Assurance

22Jan2016  
Release Date

**Medivators Inc.**  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.medivators.com](http://www.medivators.com)