



Certificate of Analysis

RAPICIDE® PA

RTU Part B

Lot Number: 028717
Expiration Date: 2019/03
Quality Control Report: 170893

Components	Specifications	Analysis
Na+	35.5 – 43.4 mEq/L	35.6
Benzotriazole	0.05% ± 0.005%	0.055
pH, (as is)	10.6 – 11.6 (as is)	11.6

This lot has been reviewed and meets specified requirements.

Quality Analyst

22/09/2017

Analysis Date

Quality Assurance

25SEP17

Release Date

Medivators Inc.

14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com