



## Certificate of Analysis

# RAPICIDE® PA

### Part A

Product Code:

**ML02-115**

Lot Number:

028718

Expiration Date:

2020/02

Components	Specifications	Analysis
H <sub>2</sub> O <sub>2</sub>	20.0 – 24.0% (w/w)	21.0 %
PAA	4.0 – 6.0 % (w/w)	5.2 %
Acetic Acid	8.0 – 10.0% (w/w)	9.1 %
pH, (as is)	0.5 – 1.1 (as is)	0.8

This lot has been reviewed and meets specified requirements.

*EPSC*

Quality Analyst

04 Sept 2018

Analysis Date

*QC*

Quality Assurance

*04 SEP 18*

Release Date

Medivators Inc.  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.medivators.com](http://www.medivators.com)

RAPICIDE® is a registered trademark of Medivators Inc.