



Certificate of Analysis

RAPICIDE® PA

Part A

Product Code: ML02-115
Lot Number: 028819
Expiration Date: 2021/01

| Components | Specifications | Analysis |
|-------------------------------|--------------------|----------|
| H ₂ O ₂ | 20.0 – 24.0% (w/w) | 23.9 % |
| PAA | 4.0 – 6.0 % (w/w) | 5.1 % |
| Acetic Acid | 8.0 – 10.0% (w/w) | 9.2 % |
| pH, (as is) | 0.5 – 1.1 (as is) | 0.5 |

This lot has been reviewed and meets specified requirements.

PSC

Quality Analyst

23-LUG-19

Analysis Date

Quality Assurance Q?

23-LUG-19

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE® is a registered trademark of Medivators Inc.