



## Certificate of Conformance

# Medline® OPA/28

### Test Strips

Part Number:	<b>MDS88OPASTP</b>
Vial Lot Number:	029115
Expiration Date:	07/17

#### Quality Control Report:

Medivators certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of  $0.50\% \pm 0.05\%$  OPA.

Medivators certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of  $0.35\% \pm 0.05\%$  OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.

Quality Assurance

1/18/2016  
Release Date

Medivators Inc.  
14605 28<sup>th</sup> Avenue N  
Minneapolis, MN 55447  
Tel: 800.444.4729  
Fax: 763.553.3387  
[www.medivators.com](http://www.medivators.com)

Medline® is a registered trademark of Medline.