



Certificate of Analysis

RAPICIDE® PA

RTU Part A

Lot Number: 029116
Expiration Date: 2018/06
Quality Control Report: 1601365

| Components | Specifications | Analysis |
|-------------------------------|-------------------|----------|
| H ₂ O ₂ | 20.0 – 24.0% | 23.4 |
| PAA | 5.0% \pm 1.0% | 5.3 |
| Acetic Acid | 8.0 – 10.0% | 9.3 |
| pH, (as is) | 0.5 - 1.1 (as is) | 0.7 |


This lot has been reviewed and meets specified requirements.



Quality Analyst

14 Dec 2016

Analysis Date



Quality Assurance PQ

15 Dec 2016

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com