



Certificate of Analysis

RAPICIDE® PA

RTU Part B

Lot Number: 029216

Expiration Date: 2018/06

Quality Control Report: _____

Components	Specifications	Analysis
Na+	35.5 – 43.4 mEq/L	<u>38.3 mEq/L</u>
Benzotriazole	0.05% \pm 0.005%	<u>0.049%</u>
pH, (as is)	10.6 – 11.6 (as is)	<u>11.2</u>

This lot has been reviewed and meets specified requirements.



Quality Analyst

20Dec2016

Analysis Date



Quality Assurance

20Dec2016

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com