



Certificate of Analysis

RAPICIDE® PA

RTU Part A

Lot Number: 029217
Expiration Date: 2019/03
Quality Control Report: 170894

Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0%	23.5
PAA	5.0% ± 1.0%	5.4
Acetic Acid	8.0 – 10.0%	8.9
pH, (as is)	0.5 - 1.1 (as is)	0.9

This lot has been reviewed and meets specified requirements.

Quality Analyst

22 Sept 2017
Analysis Date

Quality Assurance

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com