



Certificate of Conformance

RAPICIDE® Test Strip

Part Number:	ML02-0120
Lot Number:	029303
Expiration Date:	2018-07-18

I certify that the RAPICIDE test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the RAPICIDE test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

Jukan Cwt
Quality Analyst 24Aug2016
Analysis Date

Eric A. Groves
Quality Assurance 24Aug2016
Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com