



# MEDIVATORS

A Cantel Medical Company

## Certificate of Conformance

# **RAPICIDE® Test Strip**

**Part Number:**

ML02-0120

**Lot Number:**

029621

**Expiration Date:**

2018-12-14

I certify that the RAPICIDE test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the RAPICIDE test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

Andrea Herdecker  
Quality Analyst

23Jan2017

### *Analysis Date*

Jaclyn C.  
Quality Assurance

23Jan2017

*Release Date*

## Medivators Inc.

14605 28th Avenue North

Minneapolis, MN 55447 USA

Telephone: (763) 553-3300  
Fax: (763) 553-3307

[www.medivators.com](http://www.medivators.com)