



## Certificate of Analysis

# RAPICIDE® PA

### Part B

Product Code: ML02-0115  
Lot Number: 029719  
Expiration Date: 2021/01

| Components    | Specification       | Analysis    |
|---------------|---------------------|-------------|
| Sodium        | 743.0 - 909.0 mEq/L | 767.0 mEq/L |
| Benzotriazole | 0.9 – 1.1% w/w      | 1.0 %W/W    |
| pH            | 11.4 to 12.4 as is  | 11.8        |

This lot has been reviewed and meets specified requirements.

Quality Analyst

26-WG-19

Analysis Date

Quality Assurance

26-JUL-19

Release Date

Medivators Inc.  
14605 28th Avenue North  
Plymouth, Minnesota 55447  
Telephone: (763) 553-3300  
Fax (763) 553-3387

RAPICIDE® is a registered trademark of Medivators Inc.