



## Certificate of Conformance

# **RAPICIDE® Test Strip**

**Part Number:** ML02-0120

**Lot Number:** 029996

**Expiration Date:** 2019-01-17

I certify that the RAPICIDE test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the RAPICIDE test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

\_\_\_\_\_  
Quality Analyst

10Feb2017

Analysis Date

\_\_\_\_\_  
Quality Assurance

10Feb2017

Release Date

### **Medivators Inc.**

14605 28th Avenue North

Minneapolis, MN 55447 USA

Telephone: (763) 553-3300

Fax (763) 553-3387

[www.medivators.com](http://www.medivators.com)