



DOCUMENTATION

DOCUMENT
NUMBER

20100-771

REVISION: E
PAGE 7 OF 7

Quality Control report and certificate of analysis

Formula: SB-1031-R
Lot #: 02M08

Component	Concentrate Specification
Sodium	<u>2946.4</u> mEq/l $\pm 2\%$
Calcium	<u>92.1</u> mEq/l $\pm 4\%$
Potassium	<u>73.7</u> mEq/l $\pm 2\%$
Magnesium	<u>18.4</u> mEq/l $\pm 4\%$
Chloride	<u>3130.6</u> mEq/l $\pm 2\%$
Dextrose	<u>7366.0</u> mg/dl $\pm 4\%$
pH (Conc.)	<u>2.0 – 3.0</u>
pH (Dial.)	<u>6.5 – 7.4</u>
Conductivity 25°C (Conc.)	<u>149.5</u> mS/cm $\pm 5\%$
Conductivity 25°C (Dial.)	<u>13.70</u> mS/cm $\pm 5\%$

When mixed using the correct equipment and dilution with AAMI certified water Formula SB-1031-R,
Lot # 02M08, meets the above specifications.

Approved for release:

Quality Control

1/9/08
Date

RELEASED
18-May-2005