


## Quality Control Report and Certificate of Analysis

Formula: MB-330-25Lot #: 02NN07Concentrate SpecificationSodium 967.2 mEq/L  $\pm$  2 %Bicarbonate 967.2 mEq/L  $\pm$  2 %pH 7.0 - 8.0 pH unitsConductivity @ 25 °C 44.5  $\pm$  2 mS/cm

When mixed using the correct equipment and dilution with AAMI certified water MB-330-25 ,  
Lot # 02NN07 , meets the above specifications.

Approved for Release

Quality Control



Date:

02/06/07