



## Quality Control Report and Certificate of Analysis

Formula: BC-1-25  
Lot #: 02SS07

### Concentrate Specification

Sodium	<u>1187.7</u> mEq/L $\pm$ 2 %
Chloride	<u>402.6</u> mEq/L $\pm$ 2 %
Bicarbonate	<u>785.1</u> mEq/L $\pm$ 2 %
pH	<u>7.0 - 8.0</u> pH units
Conductivity @ 25 °C	<u>67</u> $\pm$ 3 mS/cm

When mixed using the correct equipment and dilution with AAMI certified water BC-1-25,  
Lot # 02SS07, meets the above specifications.

Approved for Release

Quality Control Elaine Troy Date: 05/04/07  
20600-379/D