



Quality Control Report and Certificate of Analysis

Formula: MB-330-25

Lot #: 02SS08

Concentrate Specification

Sodium 967.2 mEq/L $\pm$ 2 %

Bicarbonate 967.2 mEq/L $\pm$ 2 %

pH 7.0 - 8.0 pH units

Conductivity @ 25 °C 44.5 $\pm$ 2 mS/cm

When mixed using the correct equipment and dilution with AAMI certified water MB-330-25 ,
Lot # 02SS08 , meets the above specifications.

Approved for Release

Quality Control

A handwritten signature in black ink, appearing to read "APB", written over a horizontal line.

Date:

5-2-08