



# MEDIVATORS

A Cantel Medical Company

## Certificate of Conformance

# **RAPICIDE® Test Strip**

**Part Number:**

ML02-0120

**Lot Number:**

030288

**Expiration Date:**

2019-02-22

I certify that the RAPICIDE test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the RAPICIDE test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

Edna Danner  
Quality Analyst

Quality Analyst

28Mar2017

### Analysis Date

Andrea Heidecker  
Quality Assurance

## Quality Assurance

28Mar2017

*Release Date*

## Medivators Inc.

14605 28th Avenue North

Minneapolis, MN 55447 USA

Telephone: (763) 553-3300

Fax (763) 553-3387

www.medivators.com