



Certificate of Conformance

RAPICIDE® Test Strip

Part Number: ML02-0120

Lot Number: 030291

Expiration Date: 2019-03-29

I certify that the RAPICIDE test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the RAPICIDE test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

John A. Brown
Quality Analyst

01MAY2017
Analysis Date

Jackie Cmt
Quality Assurance

01MAY2017
Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com



TITLE

FORM, RAPICIDE GLUT TS

REV	DESCRIPTION	Effective Date	
		ORIGINATOR	ECO
D	ADD PASS/FAIL	SPW	32460
E	Update form to include test solutions, change info entry to lists/editable fields, correct spelling error, include kit and vial lot fields.	PBS	32975

Part Number	ML02-0120
Description	Rapicide Test Strips
Lot number	Vial:030291
Tested by	Erik A. Grotte
Date	01MAY2017

Note: Test form to be scanned and sent to incoming inspectors.

Vial 1

Concentration	Replicate 1	Replicate 2	Replicate 3
2.5 %	Pass	Pass	Pass
1.3 %	Pass	Pass	Pass

Vial 2

2.5 %	Pass	Pass	Pass
1.3 %	Pass	Pass	Pass

Visual Inspection

Testing Method: Unaided Eye

	Pass / Fail	Comment
Cleanliness	Pass	CLEAN
Seal of caps	Pass	SEALED
Number of test strips -0/+3	Pass	60 STRIPS PER VIAL
Adherence of test strip pad	Pass	ADHERED

Testing Solutions

Concentration	Lot# / Exp
2.5 %	852591 / 2018-09-28
1.3 %	CL17121-1 / 01JUN2017