



Certificate of Analysis


RAPICIDE® PA

Part A

Product Code: ML02-0115
Lot Number: 030316
Expiration Date: 2018/06

Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0% (w/w)	23.5 %
PAA	4.0 – 6.0 % (w/w)	5.3 %
Acetic Acid	8.0 – 10.0% (w/w)	9.3 %
pH, (as is)	0.5 – 1.1 (as is)	0.7

This lot has been reviewed and meets specified requirements.



Quality Analyst

23 Dec 2016

Analysis Date



Quality Assurance 

29 Dec 2016

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE® is a registered trademark of Medivators Inc.