



Certificate of Conformance

RAPICIDE® Test Strip

Part Number: ML02-0120

Lot Number: 030889

Expiration Date: 2019-06-07

I certify that the RAPICIDE test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the RAPICIDE test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

Quality Analyst

07JUL2017

Analysis Date

Quality Assurance

07JUL2017

Release Date

Medivators Inc.

14605 28th Avenue North
Minneapolis, MN 55447 USA

Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com