



# MEDIVATORS

A Cantel Medical Company

## Certificate of Conformance

# **RAPICIDE® Test Strip**

**Part Number:**

ML02-0120

**Lot Number:**

030893


**Expiration Date:**

2019-08-24

I certify that the RAPICIDE test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the RAPICIDE test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

  
Jennifer Bluddecke  
Quality Analyst

28Sep2017

### Analysis Date

## Quality Assurance

28Sep2017

*Release Date*

## Medivators Inc.

14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.mediators.com](http://www.mediators.com)