



Certificate of Analysis

RAPICIDE™ PA

RTU Part A

Lot Number: 031419
Expiration Date: 2021/02

Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0%	23.5%
PAA	5.0% ± 1.0%	5.3%
Acetic Acid	8.0 – 10.0%	9.0%
pH, (as is)	0.5 - 1.1 (as is)	0.6

This lot has been reviewed and meets specified requirements.



Quality Analyst

21-AUG-19

Analysis Date



Quality Assurance

21/08/19

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE™ is a trademark of Medivators Inc.