



MEDIVATORS

A Cantel Medical Company

Certificate of Conformance

RAPICIDE® Test Strip

Part Number: ML02-0120

Lot Number: 031646

Expiration Date: 2019-12-15

I certify that the RAPICIDE test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the RAPICIDE test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

Shindeh Heisbeker
Quality Analyst

Quality Analyst

27Feb2018

Analysis Date

Author Cut
Quality Assurance

Quality Assurance

27Feb2018

Release Date

Medivators Inc.

14605 28th Avenue North

Minneapolis, MN 55447 USA

Telephone: (763) 551-1111

Fax (763) 553-3387

www.medivators.com