



MEDIVATORS
A Cantel Medical Company

Certificate of Conformance

RAPICIDE® Test Strip

Part Number: ML02-0120


Lot Number: 031652

Expiration Date: 2020-04-10

I certify that the RAPICIDE test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the RAPICIDE test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.



Quality Analyst

01May2018

Analysis Date



Quality Assurance

01May2018

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com