



Certificate of Analysis

RAPICIDE[®] PA

RTU Part B

Lot Number: 033317

Expiration Date: 2019/04

Quality Control Report: 171006

| Components | Specifications | Analysis |
|---------------|---------------------|-------------------|
| Na+ | 35.5 – 43.4 mEq/L | <u>37.8 mEq/L</u> |
| Benzotriazole | 0.05% ± 0.005% | <u>0.045%</u> |
| pH, (as is) | 10.6 – 11.6 (as is) | <u>11.4</u> |

This lot has been reviewed and meets specified requirements.



Quality Analyst

19 Oct 2017
Analysis Date



Quality Assurance

23-OCT-17
Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com