



Certificate of Analysis


RAPICIDE® PA

RTU Part A

Lot Number: 033417
Expiration Date: 2019/04
Quality Control Report: 171016


Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0%	22.5%
PAA	5.0% ± 1.0%	5.7%
Acetic Acid	8.0 – 10.0%	8.7%
pH, (as is)	0.5 - 1.1 (as is)	1.1

This lot has been reviewed and meets specified requirements.


Quality Analyst

23 Oct 2017

Analysis Date


Quality Assurance

25-OCT-11

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com