



## Certificate of Analysis

# RAPICIDE® PA

## RTU Part B

Lot Number: 033517  
Expiration Date: 2019/04  
Quality Control Report: 171016

Components	Specifications	Analysis
Na+	35.5 – 43.4 mEq/L	37.5 mEq/L
Benzotriazole	0.05% ± 0.005%	0.049%
pH, (as is)	10.6 – 11.6 (as is)	11.5

This lot has been reviewed and meets specified requirements.

\_\_\_\_\_  
Quality Analyst

23 Oct 2017  
Analysis Date

\_\_\_\_\_  
Quality Assurance

25.OCT.17  
Release Date

Medivators Inc.  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
www.medivators.com