



Certificate of Analysis

RAPICIDE™ PA

RTU Part A

Lot Number: 033518
Expiration Date: 2020/04

Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0%	23.3%
PAA	5.0% ± 1.0%	5.2%
Acetic Acid	8.0 – 10.0%	9.1%
pH, (as is)	0.5 - 1.1 (as is)	0.7

This lot has been reviewed and meets specified requirements.

Quality Analyst

17 October 18

Analysis Date

Quality Assurance

23-OCT-18

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE™ is a trademark of Medivators Inc.