



## Certificate of Analysis

# RAPICIDE™ PA

### RTU Part A

Lot Number: 033718  
Expiration Date: 2020/04

Components	Specifications	Analysis
H <sub>2</sub> O <sub>2</sub>	20.0 – 24.0%	23.3%
PAA	5.0% ± 1.0%	5.7%
Acetic Acid	8.0 – 10.0%	8.7%
pH, (as is)	0.5 - 1.1 (as is)	0.8

This lot has been reviewed and meets specified requirements.

Quality Analyst

18 October 18

Analysis Date

23-OCT-18

Release Date

Medivators Inc.  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.medivators.com](http://www.medivators.com)

RAPICIDE™ is a trademark of Medivators Inc.