



Certificate of Conformance

RAPICIDE® Test Strip

Part Number: ML02-0120

Lot Number: 033761

Expiration Date: 2021-02-25

I certify that the RAPICIDE test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.


I certify that the RAPICIDE test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.


Quality Analyst

12Apr2019

Analysis Date


Quality Assurance

12Apr2019

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com



A Cantel Medical Company

Document Number:

50092-591

Rev. Level

E

TITLE

FORM, RAPICIDE GLUT TS

REV	DESCRIPTION	Effective Date	ORIGINATOR	ECO
D	ADD PASS/FAIL		SPW	32460
E	Update form to include test solutions, change info entry to lists/editable fields, correct spelling error, include kit and vial lot fields.		PBS	32975

Part Number ML02-0120
Description Rapicide Test Strips
Lot number Vial:033761 Kit:033761
Tested by Tong Vang
Date 12Apr2019

Note: Test form to be scanned and sent to incoming inspectors.

Vial 1

Concentration	Replicate 1	Replicate 2	Replicate 3
2.5 %	Pass	Pass	Pass
1.3 %	Pass	Pass	Pass

Vial 2

Concentration	Replicate 1	Replicate 2	Replicate 3
2.5 %	Pass	Pass	Pass
1.3 %	Pass	Pass	Pass

Visual Inspection

Testing Method: Unaided Eye

	Pass / Fail	Comment
Cleanliness	Pass	Clean
Seal of caps	Pass	Sealed
Number of test strips -0/+3	Pass	60 Strips per Vial
Adherence of test strip pad	Pass	Adhered

Testing Solutions

Concentration	Lot# / Exp
2.5 %	406077 / 11May2019
1.3 %	CL19101-5 / 11May2019

Tong Vang 12Apr19