



Certificate of Conformance

RAPICIDE® Test Strip

Part Number: ML02-0120

Lot Number: 034106

Expiration Date: 2021-04-03

I certify that the RAPICIDE test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the RAPICIDE test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

Quality Analyst

14May2019

Analysis Date

Quality Assurance

14May2019

Release Date

Medivators Inc.

14605 28th Avenue North
Minneapolis, MN 55447 USA

Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com



A Cantel Medical Company

Document Number:

50092-591

Rev. Level

E

TITLE

FORM, RAPICIDE GLUT TS

REV	DESCRIPTION	Effective Date	
		ORIGINATOR	ECO
D	ADD PASS/FAIL	SPW	32460
E	Update form to include test solutions, change info entry to lists/editable fields, correct spelling error, include kit and vial lot fields.	PBS	32975

Part Number ML02-0120

Description Rapicide Test Strips

Lot number Vial:034106 Kit:034106

Tested by Nick Votrobeck

Date 14May2019

Note: Test form to be scanned and sent to incoming inspectors.

Vial 1

Concentration	Replicate 1	Replicate 2	Replicate 3
2.5 %	Pass	Pass	Pass
1.3 %	Pass	Pass	Pass

Vial 2

2.5 %	Pass	Pass	Pass
1.3 %	Pass	Pass	Pass

Visual Inspection

Testing Method: Unaided Eye

	Pass / Fail	Comment
Cleanliness	Pass	Clean
Seal of caps	Pass	Sealed
Number of test strips -0/+3	Pass	60 per vial
Adherence of test strip pad	Pass	Adhered

Testing Solutions

Concentration	Lot# / Exp
2.5 %	415730 / 14Jun2019
1.3 %	CL19134-1 / 14Jun2019

N.V. 14May2019

CONFIDENTIAL AND PROPRIETARY INFORMATION OF MEDIVATORS INC. COPYING AND DISSEMINATION PROHIBITED, UNLESS AUTHORIZED BY MEDIVATORS. RETURN REQUIRED UPON DEMAND.