



Certificate of Analysis

RAPICIDE™ PA

RTU Part B


Lot Number: 0035719

Expiration Date: 2021/03

Components	Specifications	Analysis
Na+	35.5 – 43.4 mEq/L	<u>37.1mEq/L</u>
Benzotriazole	0.05% ± 0.005%	<u>0.054%</u>
pH, (as is)	10.6 – 11.6 (as is)	<u>11.6</u>

This lot has been reviewed and meets specified requirements.

 23-SEP-13
Quality Analyst Analysis Date

 25-SEP-13
Quality Assurance Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE™ is a trademark of Medivators Inc.