



Certificate of Analysis

RAPICIDE™ PA

RTU Part A

Lot Number: 036718
Expiration Date: 2020/05


Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0%	<u>23.9%</u>
PAA	5.0% ± 1.0%	<u>5.6%</u>
Acetic Acid	8.0 – 10.0%	<u>8.7%</u>
pH, (as is)	0.5 - 1.1 (as is)	<u>0.7</u>

This lot has been reviewed and meets specified requirements.


Quality Analyst


13-NOV-18
Analysis Date


Quality Assurance


18-nov-18
Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE™ is a trademark of Medivators Inc.