



## Certificate of Analysis

# RAPICIDE® PA

### Part B

Product Code: ML02-0115  
Lot Number: 036719  
Expiration Date: 2021/03

Components	Specification	Analysis
Sodium	743.0 - 909.0 mEq/L	745.0 mEq/L
Benzotriazole	0.9 – 1.1% w/w	1.0 %w/w
pH	11.4 to 12.4 as is	11.6

This lot has been reviewed and meets specified requirements.

*PSC*

Quality Analyst

*Q*

Quality Assurance

*26-SET-19*

Analysis Date

*30-SEP-19*

Release Date

Medivators Inc.  
14605 28th Avenue North  
Plymouth, Minnesota 55447  
Telephone: (763) 553-3300  
Fax (763) 553-3387

RAPICIDE® is a registered trademark of Medivators Inc.