



## Certificate of Analysis

# RAPICIDE® PA

### RTU Part A

Lot Number: 037517  
Expiration Date: 2019/05  
Quality Control Report: 171123

Components	Specifications	Analysis
H <sub>2</sub> O <sub>2</sub>	20.0 – 24.0%	22.7%
PAA	5.0% ± 1.0%	5.8%
Acetic Acid	8.0 – 10.0%	8.6%
pH, (as is)	0.5 - 1.1 (as is)	1.1

This lot has been reviewed and meets specified requirements.

*201123*  
Quality Analyst

*20 November 17*  
Analysis Date

*22 Nov 17*  
Quality Assurance

Release Date

Medivators Inc.  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.medivators.com](http://www.medivators.com)