

*Certificate of Conformance*

**RAPICIDE® OPA-28 Test Strips**

**Part Number:**

**ML02-0137**

**Vial Lot Number:**

038116

**Expiration Date:**

2018-07-31

**Quality Control Report:**

MEDIVATORS certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of  $0.50\% \pm 0.05\%$  OPA.

MEDIVATORS certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of  $0.35\% \pm 0.05\%$  OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.



Quality Assurance

1/26/2017

Release Date

**Medivators Inc.**  
14605 28<sup>th</sup> Avenue N  
Minneapolis, MN 55447  
Tel: 800.444.4729  
Fax: 763.553.3387  
[www.medivators.com](http://www.medivators.com)

RAPICIDE® is a registered trademark of Medivators Inc.