

*Certificate of Conformance*

**RAPICIDE® OPA/28 Test Strips**

|                         |                  |
|-------------------------|------------------|
| <b>Part Number:</b>     | <b>ML02-0136</b> |
| <b>Lot Number:</b>      | 038117           |
| <b>Vial Lot Number:</b> | 049117           |
| <b>Expiration Date:</b> | 2019-11-30       |

**Quality Control Report:**

MEDIVATORS certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of 0.50%  $\pm$  0.05% OPA.

MEDIVATORS certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of 0.35%  $\pm$  0.05% OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.

  
Quality Assurance

1/23/2018  
Release Date

**MEDIVATORS Inc.**  
14605 28<sup>th</sup> Avenue N  
Minneapolis, MN 55447  
Tel: 800.444.4729  
Fax: 763.553.3387  
www.medivators.com