

Certificate of Conformance

McKESSON[®] OPA/28 Test Strips

Part Number:	73-OPA28B
Vial Lot Number:	039017
Expiration Date:	2018-09-30

Quality Control Report:

MEDIVATORS certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of 0.50% $\pm$ 0.05% OPA.

MEDIVATORS certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of 0.35% $\pm$ 0.05% OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.



Quality Assurance

2/17/2017
Release Date

Medivators Inc.
14605 28th Avenue N
Minneapolis, MN 55447
Tel: 800.444.4729
Fax: 763.553.3387
www.medivators.com

McKESSON[®] is a registered trademark of McKesson Corporation.