

Certificate of Conformance

RAPICIDE® OPA-28 Test Strips

Part Number:

ML02-0137

Vial Lot Number:

039017

Expiration Date:

2018-09-30

Quality Control Report:

MEDIVATORS certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of 0.50% $\pm$ 0.05% OPA.

MEDIVATORS certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of 0.35% $\pm$ 0.05% OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.



Quality Assurance

1/26/2017

Release Date

Medivators Inc.

14605 28th Avenue N
Minneapolis, MN 55447
Tel: 800.444.4729
Fax: 763.553.3387
www.medivators.com

RAPICIDE® is a registered trademark of Medivators Inc.