



## Certificate of Analysis

# RAPICIDE™ PA

## RTU Part B


Lot Number: 0039019  
Expiration Date: 2021/04

Components	Specifications	Analysis
Na+	35.5 – 43.4 mEq/L	<u>35.5mEq/L</u>
Benzotriazole	0.05% $\pm$ 0.005%	<u>0.053%</u>
pH, (as is)	10.6 – 11.6 (as is)	<u>11.6</u>

This lot has been reviewed and meets specified requirements.

  
Quality Analyst

15-OCT-19  
Analysis Date

  
Quality Assurance

17-OCT-19  
Release Date

Medivators Inc.  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.medivators.com](http://www.medivators.com)

RAPICIDE™ is a trademark of Medivators Inc.