



MINNTECH®

DOCUMENTATION

DOCUMENT
NUMBER

20100-771

REVISION: E

PAGE 7 OF 7

Quality Control report and certificate of analysis

Formula: SB-1031-R
Lot #: 03M08

<u>Component</u>	<u>Concentrate Specification</u>
Sodium	2946.4 mEq/l \pm 2%
Calcium	92.1 mEq/l \pm 4%
Potassium	73.7 mEq/l \pm 2%
Magnesium	18.4 mEq/l \pm 4%
Chloride	3130.6 mEq/l \pm 2%
Dextrose	7366.0 mg/dl \pm 4%
pH (Conc.)	2.0 – 3.0
pH (Dial.)	6.5 – 7.4
Conductivity 25°C (Conc.)	149.5 mS/cm \pm 5%
Conductivity 25°C (Dial.)	13.70 mS/cm \pm 5%

When mixed using the correct equipment and dilution with AAMI certified water Formula SB-1031-R, Lot # 03M08, meets the above specifications.

Approved for release:

Elaine Trojan
Quality Control

01/23/08
Date

RELEASED
18-May-2005

THIS DOCUMENT AND ALL INFORMATION CONTAINED HEREIN IS THE PROPERTY OF MINNTECH CORPORATION AND IS PROPRIETARY AND CONFIDENTIAL. THIS CONFIDENTIAL INFORMATION MUST NOT BE MADE PUBLIC, DISCLOSED TO ANY THIRD PARTY OR COPIED UNLESS AUTHORIZED BY AN OFFICER OF MINNTECH, AND IS SUBJECT TO RETURN UPON DEMAND